



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: JAN 4, 2011 INITIALS: DJZ OWNER: CITY OF TACOMA
DISTRICT: ADDRESS: COUNTY:
LOCATION: TYPE PORT. WHEEL SCALE ACCURACY CLASS: IIII
MAKE: HAENNI ID. NO: 21965 CAPACITY: 20000 lbs. SECTIONS:
MODEL: 101 REG. ELEMENT: DIAL PLATFORM 15"X26"
MINIMUM GRADUATIONS: 50 lbs.


LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
CENTER	5000	5000	0				
CENTER	10000	10000	0				
CENTER	15000	15000	0				
CENTER	20000	20000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report and Certificate of Calibration",
Certification number(s): 36293-W-02

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy

Signed by: Dan J. LaFreniere Date: Jan 20, 2011

Print name: Dan LaFreniere



State of Washington
County of Thurston
Signed and sworn to (or affirmed before
me on Jan 20, 2011 by

Brenda A Biscay
(Signature of notary public)
Brenda A. Biscay
Printed Name of Notary
My appointment expires 3.12.12